

Research Infrastructure Quality Assurance

GAW Report No. 253

# International Comparison of Dobson Spectrophotometers

Pretoria, Gauteng Province, South Africa  
7–18 October 2019

WEATHER CLIMATE WATER



WORLD  
METEOROLOGICAL  
ORGANIZATION



GLOBAL  
ATMOSPHERE  
WATCH



Research Infrastructure Quality Assurance

GAW Report No. 253

# International Comparison of Dobson Spectrophotometers

Pretoria, Gauteng Province, South Africa  
7-18 October 2019

Prepared by U. Köhler, G.J.R. Coetzee, G. McConville



**© World Meteorological Organization 2020**

The right of publication in print, electronic and any other form and in any language is reserved by WMO. Short extracts from WMO publications may be reproduced without authorization, provided that the complete source is clearly indicated. Editorial correspondence and requests to publish, reproduce or translate this publication in part or in whole should be addressed to:

Chairperson, Publications Board  
World Meteorological Organization (WMO)  
7 bis, avenue de la Paix  
P.O. Box 2300  
CH-1211 Geneva 2, Switzerland

Tel.: +41 (0) 22 730 84 03  
Fax: +41 (0) 22 730 80 40  
E-mail: [Publications@wmo.int](mailto:Publications@wmo.int)

**NOTE**

The designations employed in WMO publications and the presentation of material in this publication do not imply the expression of any opinion whatsoever on the part of WMO concerning the legal status of any country, territory, city or area, or of its authorities, or concerning the delimitation of its frontiers or boundaries.

The mention of specific companies or products does not imply that they are endorsed or recommended by WMO in preference to others of a similar nature which are not mentioned or advertised.

The findings, interpretations and conclusions expressed in WMO publications with named authors are those of the authors alone and do not necessarily reflect those of WMO or its Members.

This publication has been issued without formal editing.

## CONTENTS

GROUP PHOTO .....	iv
BACKGROUND .....	1
PARTICIPATION AND PROCEEDINGS .....	1
FINAL RESULTS AND RECOMMENDATIONS .....	3
Technical Reports .....	3
General Recommendations.....	3
Appendix A – List of participants .....	4
Appendix B – Individual Instrument Reports .....	6
Instrument D015 - Botswana (Maun) .....	8
Instrument D018 - Kenya (Nairobi) .....	12
Instrument D035 - South Africa (Stellenbosch) .....	17
Instrument D064 - Germany (MOHp, European regional standard).....	22
Instrument D089 - South Africa (Irene).....	25
Instrument D112 - India (New Delhi) .....	31
Instrument D132 - South Africa (Springbok) .....	36



**Group photo of the participants of the IRENE2019 DIC in South Africa**

## BACKGROUND

The World Meteorological Organization (WMO) Secretariat's (Global Atmosphere Watch (GAW) programme) and the South African Weather Service (SAWS) with close cooperation and assistance of the National German Meteorological Service (DWD) Regional Dobson Calibration Centre (RDCC-E) at the Meteorological Observatory Hohenpeissenberg (MOHp) and the USA National Oceanic and Atmospheric Administration's Earth Systems Research Laboratory's Global Monitoring Division (NOAA/ESRL/GMD, World Dobson Calibration Centre WDCC) organized the DIC IRENE2019. IRENE2019 was a campaign to maintain the network of the Dobson ozone spectrophotometers operated in the southern African region, and to serve as an assurance of the quality of the total ozone datasets created at the regional stations. This action is a fulfilment of WMO/GAW/QC requirements for monitoring of atmospheric total ozone. The Dobson spectrophotometers of countries from the northern African region were already calibrated in advance during the WMO European DIC at INTA, El Arenosillo in September 2017 and an internal Egyptian campaign at Hurgada in September 2019 (Egyp2019).

The South African Weather Service, under the auspices of the World Meteorological Organization (WMO) hosted the 4<sup>th</sup> African Regional Dobson Spectrophotometer Intercomparison DIC IRENE2019 meeting from 7 to 18 October 2019, under the supervision of Mr Gerrie Coetsee. Mr Ulf Köhler took the position of technical director for the scientific soundness of the work undertaken.

Funding for this international event was provided from the UNEP Vienna Convention Trust Fund (VCTF) in support for WMO to assist countries in their efforts for monitoring and the protection of the Ozone Layer.

The participation of the German team (Mr U. Köhler and Mr M. Heinen) was completely funded by the DWD. NOAA funded Mr G. McConville's participation except for travel that came from the VCTF. South Africa funded their own participations of three Dobson's and participating resource staff.

## PARTICIPATION AND PROCEEDINGS

The technical and scientific aspects of the IRENE2019 event were supervised by Mr Ulf Köhler from the RDCC-E and by Mr Glen McConville from the WDCC, who supplied the reference standards D064 (Europe) and D065 (Travelling World Standard) and much appreciated technical equipment.

Eleven specialists from six countries participated at the DIC and are listed Appendix A. The Primary World Standard Dobson (D065) from NOAA and the Regional Standard Dobson (D064) were used as the references for the instruments noted below. Additionally Dr Siddharta Singh from the India Meteorological Department also participated with the Indian instrument D112. The planned calibration of the Shimadzu 5703 from Nigeria could not be performed, as the instrument arrived too late due to transportation/customs problems and also the participant could not obtain his visa in time. The participation of the instrument from the Seychelles (D057) was already cancelled in preparation of the campaign. Difficulties were also experienced as the two Botswana participants could only join the event on the last three days

of activities. The following national Dobson spectrophotometers were inspected, adjusted, and compared at the IRENE2019:

<i>Dobson No.</i>	<i>Country</i>	<i>Station</i>	<i>Purpose</i>
15	Botswana	Maun	regular calibration
18	Kenya	Nairobi	regular calibration
35	South Africa	Stellenbosch	regular calibration
64	Germany	Hohenpeissenberg	standard & calibration check with D065
65	USA	Boulder	standard
89	South Africa	Irene	regular calibration
112	India	New Delhi	calibration to resume operation
132	South Africa	Springbok	regular calibration

The main tasks of IRENE2019 were:

- 1) To evaluate each instrument calibration and the existing total ozone datasets.
- 2) Repair, clean and improve the instrument as needed.
- 3) Redefine a new calibration levels (where applicable) for the future measurements at the home station before returning the instruments in good operating order.

Besides the stated goals of Dobson intercomparison campaigns, such an event also provides a forum for sharing of information between observing programmes, and serves as a venue for training. These goals were successfully attained during this campaign. The experts from the calibration centres had enough time and good opportunities to train the less experienced participants. Educational video clips, how to do measurements and tests, are recommended for training purposes and can be watched under:

<http://www.o3soft.eu/dobsonweb/instrument.html>

As each instrument arrived, it and its accessories were inspected. At this stage no repairs, except those needed to make the instrument operate were made. The instrument was then operated in the condition as it would have been at the home station. Once a successful initial Intercomparison (IC) is made, those results then dictate whether significant optical, mechanical, or electronic repairs were needed. If needed, a final Intercomparison (FC) is performed to define a new calibration.

Other tasks included cleaning and/or repair/replacement of optical, electronic and mechanical parts followed by discharge lamp tests to create new Q-tables and an optical wedge calibration to define new R-G-tables, if needed. A series of lamp calibrations were also done on a daily basis.

The Regional Standard for Europe, D064 (DWD, Hohenpeissenberg, Germany), was compared with the World Standard D065 (NOAA, Boulder, USA) in order to check and confirm its calibration level. This event should be done every two to three years to guarantee the correct transfer of the calibration level of the World Standards D065 and D083 into the WMO RA VI Europe Dobson network.



## FINAL RESULTS AND RECOMMENDATIONS

### Technical Reports

The official technical report consisting of the detailed worked performed and assessment of each individual instrument is presented in Appendix B. All the participating instruments listed above are now well calibrated and in good operational condition. The revised IC Objectives were met satisfactorily under challenging circumstances. The calibration level of the regional Standard D064 was successfully confirmed.

### General Recommendations

The Dobson/Shimadzu instruments and staff from Seychelles and Nigeria were not able to participate. These programmes should be assisted to be able to take part in other scheduled IC events, if the responsible institutes evince high interest in continuation of their ozone monitoring programme using Dobson/Shimadzu instruments. Contact has been established with the responsible persons on the Seychelles, how their programme can be resumed (with a repaired/refurbished Dobson or with a new instrument) with the support of UNEP and or WMO. It is also recommended that the Dobson and participant of Nigeria attend another future IC campaign if this can be accommodated.

One important finding of the campaign was the fact, that the regular WMO recommendation for calibration intervals of no longer than five to six years is absolutely justified. As the last regular DIC took place ten years ago in 2009, some of the instruments show significantly larger deviations than  $\pm 1\%$  from the standard Dobsons. Furthermore it is helpful to use such campaigns for training purposes of the stations' Dobson staff, to improve the performance of the instruments and to avoid uncertain or even poor data quality because of not correct procedures of measurements and regular tests. The above mentioned video clips will surely support this goal.

The able assistance of Mr Glen McConville, Mr Ulf Köhler and Mr Michael Heinen in both the technical achievements of the intercomparison and the scientific assessment in this report is acknowledged, with appreciation.

**SAWS/WMO/GAW – IRENE2019**  
**International Comparison of Dobson Spectrophotometers (DIC)**  
**Irene Technical Centre, Pretoria, Gauteng Province,**  
**South Africa, 7-18 October 2019**

**List of Participants**

**Botswana**

Mr Karabo Molefe  
Botswana Meteorological Services,  
P.O. Box 458 Maun, Meteorological Office Maun,  
Maun airport  
Tel.: +267 6860429 or 6800595  
Email: krmolefe@gov.bw

Mrs Goitsewang Salethake  
Botswana Meteorological Services  
Plot 54216, Corner Maaloso/ Metsimotlhaba Rd, Gaborone Village  
P.O. Box 10100, Gaborone  
Tel.: +267 3612200  
Fax: +267 3956282  
Email: gsalethake@gov.bw

**Germany**

Mr Ulf Köhler  
Deutscher Wetterdienst  
Meteorological Observatory Hohenpeißenberg  
Albin-Schwaiger- Weg, 10  
82383 Hohenpeißenberg  
Tel.: 0049 69 8062 9771  
Email: Ulf.Koehler@dwd.de

Mr Michael Heinen  
Deutscher Wetterdienst  
Meteorological Observatory Hohenpeißenberg  
Albin-Schwaiger- Weg, 10  
82383 Hohenpeißenberg  
Email: Michael.Heinen@dwd.de

**India**

Dr Siddhartha Singh  
Email: siddhartha.singh74@imd.gov.in  
India Meteorological Department  
Ozone & Air Quality Unit  
Environment Monitoring & Research Centre  
New Delhi – 110003  
Tel.: 0091 114 3824235

**Kenya**

Mrs Annet Nyaga  
Kenya Meteorological Department  
P. O. Box 30259,  
00100 GPO Nairobi  
Email: nyagaannet@gmail.com

**South Africa**

Mr Gerrie J R Coetzee  
Email: gerrie.coetzee@weathersa.co.za  
South African Weather Service  
Eco Glades block 1b, Eco Park  
Cnr Olievenhoutbosch and Ribbon Grass Streets  
Centurion, 0157  
Pretoria 001  
Tel: +27 12 367 6048  
Fax: +27 12 367 6189

Mr Daniel van der Spuy  
South African Weather Service, C/O CCSIR,  
11 Cilliers road  
Stellenbosch  
Email: Danie.vanderSpuy@weathersa.co.za

Mr Gerald Gerhardus Meyer  
South African Weather Service  
Springbok, 8240  
Cell: 0636954879  
Tel.: 02771 22961  
Email: Gerald.Meyer@weathersa.co.za

Ms Tshidi Machinini  
South African Weather Service  
Eco Glades block 1b, Eco Park  
Cnr Olievenhoutbosch and Ribbon Grass Streets  
Centurion, 0157, Pretoria 001  
Email Tshidi.Machinini@weathersa.co.za

**USA**

Mr Glen McConville  
U.S. Department of Commerce  
National Oceanic and Atmospheric Administration  
Earth System Research Laboratory, Global Monitoring Division  
Mail Code R/GMD-1  
325 Broadway  
Boulder, Colorado 80305  
Telephone: 1 303 497 3989  
Email: glen.mcconville@noaa.gov

**SAWS/WMO/GAW – IRENE2019**  
**International Comparison of Dobson Spectrophotometers (DIC)**  
**Irene Technical Centre, Pretoria, Gauteng Province,**  
**South Africa, 7-18 October 2019**

**Individual Instrument Reports**

**Intercomparison Results**  
**Iren2019 from October 7 to 18, 2019 at Irene**

**Instrument D015**  
**Botswana (Maun)**

<b>Initial calibration (18.10.2019):</b>						
	<b>Date:</b>	<b>Comment:</b>				
G-Tables	8.3.94	after wedge-cal. on March. 8, 1994, location?				
N-Tables	26.11.09	after FC on November 26, 2009 at Irene2009				
<b>Corrections to N-Tables</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
from SL-Test	-0.60	-0.90	-0.70	0.10	-0.20	
from Comparison	1.40	-0.07	-0.24	1.64	0.17	
Sum	0.80	-0.97	-0.94	1.74	-0.03	
<b>Comments:</b> Initial = final calibration as only one intercomparison was possible poor agreement in AD, data noisy, no reprocessing as not operational						
<b>Final Calibration (18.10.2019):</b>						
	<b>Date:</b>	<b>Comment:</b>				
G-Tables	8.3.94	after wedge-cal. on March. 8, 1994, location?				
N-Table (new)	18.10.19	after FC on October 18, 2019 at Irene2019				
<b>Corrections to:</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
old N-Table	0.80	-0.97	-0.94	1.74	-0.03	
<b>Comments:</b> Initial = final calibration as only one intercomparison was possible new R-N-table created, should be used for future observations						
<b>Reference Standard Lamp Data:</b>		<b>Date:</b>		18.10.19		
for new R-N-table						
Lamp No.	A		C		D	
	R	N	R	N	R	N
15Q1	20.30	10.68	20.50	15.79	20.10	17.61
15Q2	20.60	10.98	20.30	15.59	20.30	17.82
<b>Q-Table:</b> Check new Q-table from discharge lamp test October 17 after return by doing an HG-test and correct if necessary according the procedure described in Komhyr's old or Bob's new Dobson manual						

## **Instrument D015 Botswana (Maun)**

### **Original calibration data**

N-tables from 26 November 2009 based on DSGQP-comparison with D083 on 26 November 2009 at Irene2009, G-tables from 8 March 1994, location unknown.  
Reference Standard Lamp Values for lamps 15Q1 and 15Q2  
(Lamp tests results used in data processing at home station)?

### **Introductory remarks**

Instrument not operational in the past years. Arrival very late, thus only one intercomparison possible and no additional work could be done.

### **Initial calibration results**

(Adjustments based on the results with Standard Lamp tests included)  
18 October 2019:

**d\_Na: -1.40 d\_Nc: -0.07 d\_Nd: -0.24 d\_Nad: 1.64 d\_Ncd: 0.17**

The d\_Nad value implies an average **-2.3% error** in calculated ozone value,  $\mu = 1.15$  to  $2.5$ , Total Ozone = 300 Dobson Units. Data noisy; no reprocessing necessary, as instrument not operational.

### **Optical, mechanical and electrical work performed**

- Electric/Electronics: new US-type (original).
- Optical check: Q-plates, Ls and Ps dusty, Ms with scratches (M1) and cloudy (M2), wedge and cobalt filter O.K., prism sundirector dusty and head reversed and loose.
- Symmetry test: Not done.
- Measurement of slit widths and parallelism with microscope: Not done.
- Shutter motor: Toothbelt drive, 815 rpm.
- PMT vertical position test: Not done, Focus L1: Not done.
- Optics: Optics cleaned and loose sun director head fixed.
- Discharge lamp: Performed on 17 October 2019, new Q-table created.
- Wedge calibration: Not done.

### **Final intercomparison**

18 October 2019

IC and FC on the same day, new R-N-tables created; **highest difference against the standard ADDSGQP observations of D064 in  $\mu$  range 1.15 to 3.2 was 0.9 % in total ozone, moderate  $\mu$ -dependency. CD-observations with large  $\mu$ -dependence, but not used.**

### **Recommendations/comments**


- The results of the initial/final calibration are not good, but no re-calculation necessary as instrument not operational. Final data can be used for:
- New calibration status with new R-N-tables defined based on old R-G/N-tables incl. determination of SL-reference values from corresponding tests (s. table with results).
- **Correction of the used-Q-tables at Irene derived from results of the Hg-Test immediately after return to station according the Dobson instructions.**

- Regular test (monthly SL and HG, at least annual Symmetry Test) and cleaning of GQP/Sundirector.
- **New absorption coefficients will officially be introduced in the near future: reprocessing of all data necessary. Information will be given. Question of the future use of effective absorption coefficients (derived from optical characterization) not finally clarified.**


Hohenpeissenberg, 19.12.2019

### Difference N\_A-C-D to Reference Instrument D064 Initial/Final Calibration on 18 October 2019

N(D064)-N(D015) vs. Mue for D064 & D015 Irene, 18. October 2019  
Initial/Final Calibration original data



N(D064)-N(D015) vs. Mue for D064 & D015 Irene, 18. October 2019  
Initial/Final Calibration adjusted data





**Intercomparison Results**  
**Iren2019 from October 7 to 18, 2019 at Irene**

**Instrument D018**  
**Kenya (Nairobi)**

<b>Initial calibration (12.10.2019):</b>						
	<b>Date:</b>	<b>Comment:</b>				
<b>G-Tables</b>	<b>11.4.05</b>	after wedge-cal. on April 11, 2005 at MOHp2005				
<b>N-Tables</b>	<b>14.4.05</b>	after FC on April 14, 2005 with new R-G at				
<b>Corrections to N-Tables</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
from SL-Test	0.70	0.90	1.00	-0.30	-0.10	
from Comparison	-0.32	-0.34	-1.02	0.70	0.68	
<b>Sum</b>	<b>0.38</b>	<b>0.56</b>	<b>-0.02</b>	<b>0.40</b>	<b>0.58</b>	
<b>Comments:</b> Good agreement in AD, larger differences in CD and single D No data reprocessing necessary, as CD not used						
<b>Final Calibration (18.10.2019):</b>						
	<b>Date:</b>	<b>Comment:</b>				
<b>G-Tables</b>	<b>11.4.05</b>	after wedge-cal. on April 11, 2005 at MOHp2005				
<b>N-Table (new)</b>	<b>18.10.19</b>	after FC on Oct. 18, 2019 with old R-N at Irene2019				
<b>Corrections to:</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
old N-Table	0.70	0.90	0.30	0.40	0.60	
<b>Comments:</b> Good agreement in AD and CD New RN-tables created for future data processing						
<b>Reference Standard Lamp Data:</b>		<b>Date:</b>		<b>18.10.19</b>		
for new R-N-table						
Lamp No.	A		C		D	
	R	N	R	N	R	N
18V	61.20	13.50	62.50	17.45	62.90	19.00
18W	60.70	13.00	62.20	17.10	62.70	18.80
<b>Q-Table:</b> Check original Q-table in use at station by doing an HG-test and correct if necessary according the procedure described in Komhyr's old or Bob's new Dobson manual						

## **Instrument D018 Kenya (Nairobi)**

### **Original calibration data**

N-tables from 14 April 2005 based on DSGQP-comparison with D064 on 14 April 2005, MOHp, G-tables from 11 April 2005.  
Reference Standard Lamp Values for lamps 18V and 18W.  
Lamp tests results used in data processing at home station.

### **Introductory remarks**

The sometimes large deviations in the SL-tests were found not to be realistic. In these cases the SL-tests were probably made without ground quartz plate (GQP). Corresponding tests with and without GQP at Irene confirmed this assumption. **Thus check the "old" data: if corrections have been applied based on the wrong SI-tests without GQP, these data have to be reprocessed.**

### **Initial calibration results**

(Adjustments based on the results with Standard Lamp tests included)  
12 October 2019:

**d\_Na: -0.32 d\_Nc: -0.34 d\_Nd: -1.02 d\_Nad: 0.70 d\_Ncd: 0.68**

The d\_Nad value implies an average **+1.00% error** in calculated ozone value,  $\mu=1.15$  to 2.5, Total Ozone = 300 Dobson Units. Good agreement in AD, larger differences in CD and single D; no data reprocessing necessary, as CD not used.

### **Optical, mechanical and electronical work performed**

- Electric/Electronics: New US type MOHp-modified.
- Optical check: Sun director, Q-plates, Lenses and Prisms slightly dusty, Mirrors, wedge plates and cobalt filter O.K.
- Symmetry test: Results acceptable.
- Measurement of slit widths and parallelism with microscope: Not done.
- Shutter motor: Belt drive, 865 rpm.
- PMT vertical position test: Not done, Focus L1: Not done.
- Optics: Dusty optics cleaned.
- Discharge lamp: Not done.
- Wedge calibration: Not done.

### **Final intercomparison**

17 and 18 October 2019

Data of two final calibration processed, data of 17 October too noisy, therefore only 18 October used FC. **New R-N-tables created and should be used for future data processing; highest difference against the standard ADDSGQP observations of D064 in  $\mu$  range 1.15 to 3.2 was -0.65 % in total ozone, small  $\mu$ -dependency.**

### **Recommendations/comments**

- The results of the initial calibration are good, therefore no re-calculation necessary (**only if wrong SL-corrections were applied**), especially as only AD used for data evaluation.

- New calibration status with new R-N-tables defined based on old R-N-tables incl. determination of SL-reference values from corresponding tests (s. table with results).
- **Correction of the used-Q-tables at Irene derived from results of the Hg-Test immediately after return to station according the Dobson instructions.**
- Regular test (monthly SL and HG, at least annual Symmetry Test) and cleaning of GQP/Sundirector.
- **New absorption coefficients will officially be introduced in the near future: reprocessing of all data necessary. Information will be given. Question of the future use of effective absorption coefficients (derived from optical characterization) not finally clarified.**


Hohenpeissenberg, 20.12.2019

### Difference N\_A-C-D to Reference Instrument D064

#### Initial Calibration on 12 October 2019


N(D064)-N(D018) vs. Mue for D064 & D018 Irene, 12. October 2019

Initial Calibration original data





N(D064)-N(D018) vs. Mue for D064 & D018 Irene, 12. October 2019

Initial Calibration adjusted data



### Difference N\_A-C-D to Reference Instrument D064 Final Calibration on 18 October 2019



**Intercomparison Results**  
**Iren2019 from October 7 to 18, 2019 at Irene**

**Instrument D035**  
**South Africa (Stellenbosch)**

<b>Initial calibration (09.10.2019):</b>						
	<b>Date:</b>	<b>Comment:</b>				
G-Tables	15.7.04	after wedge-cal. on July. 15, 2004, at MOHp 2004				
N-Tables	27.10.09	after FC on Oct. 27, 2009, at Irene2009				
<b>Corrections to N-Tables</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
from SL-Test	0.20	0.50	0.40	-0.20	0.10	
from Comparison	0.44	0.66	1.15	-0.71	-0.49	
Sum	0.64	1.16	1.55	-0.91	-0.39	
<b>Comments:</b> 1. IC on October 9, 2019, with old RN-tables: acceptable agreement in AD Just within 1%-limit, no data reprocessing recommended 2. IC on October 12, 2019, not used as data too noisy						
<b>Final Calibration (18.10.2019):</b>						
	<b>Date:</b>	<b>Comment:</b>				
G-Tables	15.7.04	after wedge-cal. on July. 15, 2004, at MOHp 2004				
N-Table (new)	18.10.19	after FC on Oct. 18, 2019, at Irene2019				
<b>Corrections to:</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
old N-Table	3.10	3.50	4.10	-1.00	-0.60	
<b>Comments:</b> 2.FC on October18 with good result in AD at not too high sun, Data should not be used at Mue-values < 1.4 New RN-tables created for future data processing						
<b>Reference Standard Lamp Data:</b>			<b>Date:</b>		18.10.19	
for new R-N-table						
Lamp No.	A		C		D	
	R	N	R	N	R	N
35Q2	33.07	9.08	36.30	14.06	39.43	15.78
35Q3	33.60	9.59	36.43	14.18	39.73	16.07
<b>Q-Table:</b> Check new Q-table created after discharge lamp test on 18.10. at station by doing an HG-test and correct if necessary according the procedure described in Komhyr's old or Bob's new Dobson manual						

## **Instrument D035 South Africa (Stellenbosch)**

### **Original calibration data**

N-tables from 27 October 2009, based on DSGQP-comparison with D083 on 27 October 2009 at Irene, G-tables from 15 July 2004 at MOHp.

Reference Standard Lamp Values for lamps 35Q2 and 35Q3.

Lamp tests results used in data processing at home station.

### **Introductory remarks**

Instrument operational at Stellenbosch. Second initial calibration on 12 October and first final calibration on 17 October were not used as data too noisy.

### **Initial calibration results**

(Adjustments based on the results with Standard Lamp tests included)

9 October 2019:

**d\_Na: 0.44 d\_Nc: 0.66 d\_Nd: 1.15 d\_Nad: -0.71 d\_Ncd: -0.49**

The d\_Nad value implies an average **+1.0% error** in calculated ozone value,  $\mu=1.15$  to 2.5, Total Ozone = 300 Dobson Units. Acceptable agreement with moderate mu-dependance in AD, no data reprocessing recommended.

### **Optical, mechanical and electrical work performed**

- Electric/Electronics: New US type MOHp-modified.
- Optical check: Lenses, Q-plates and wedge dusty, all other optics O.K.
- Symmetry test: Done on 8 and 18 October with good results.
- Measurement of slit widths and parallelism with microscope: Not done.
- Shutter motor: Friction drive, speed with 750 rpm a little bit low.
- PMT vertical position test: Not done, Focus L1: Not done.
- Optics: Lenses, Q-plates and wedge cleaned.
- Further work: Function of wedge sled stiff, therefore greased.
- Discharge lamp: Done on 18 October, new Q-table created, which should be used back at station.
- Wedge calibration: Done in October with bad results because of fingerprints after first cleaning, cleaned again and C-wedge cal repeated, which shows no difference to wedge calibration in 2004.

### **Final intercomparison**

17 and 18 October 2019

Data of two final calibration processed, data of October 17 too noisy, therefore only 18 October used as FC. **New R-N-tables created and should be used for future data processing; highest difference against the standard ADDSGQP observations of D064 in mu range 1.15 to 3.2 was 1.16 % in total ozone at high sun, thus data**

**with  $\mu_e < 1.4$  should not be used. Reason for too high ozone values at high sun (in contrast to 1.IC) not known.**

#### **Recommendations/comments**

- The results of the initial calibration are satisfactory, therefore no re-calculation necessary, especially as only AD used for data evaluation.
- New calibration status with new R-N-tables defined based on old R-N-tables incl. determination of SL-reference values from corresponding tests (s. table with results), data at high sun ( $\mu_e < 1.4$ ) should not be used.
- **Correction of new Irene-Q-table (after discharge lamp test on 18 October 2019), from results of the Hg-Test immediately after return to station, if necessary.**
- Regular test (monthly SL and HG, at least annual Symmetry Test) and cleaning of GQP/Sundirector.
- **New absorption coefficients will officially be introduced in the near future: reprocessing of all data necessary. Information will be given. Question of the future use of effective absorption coefficients (derived from optical characterization) not finally clarified.**


Hohenpeissenberg, 10.01.2020




## Difference N\_A-C-D to Reference Instrument D064

### Initial Calibration on 9 October 2019

N(D064)-N(D035) vs. Mue for D064 & D035 Irene, 9. October 2019  
Initial Calibration original data - truncated & edit



N(D064)-N(D035) vs. Mue for D064 & D035 Irene, 9. October 2019  
Initial Calibration adjusted data - truncated & edit




### Difference N\_A-C-D to Reference Instrument D064

#### Final Calibration on October 18, 2019


N(D064)-N(D035) vs. Mue for D064 & D035 Irene, 18. October 2019

Final Calibration original data



N(D064)-N(D035) vs. Mue for D064 & D035 Irene, 18. October 2019

Final Calibration adjusted data



**Intercomparison Results**  
**Iren2019 from October 7 to 18, 2019 at Irene**

**Instrument D064**  
**Germany (Hohenpeissenberg)**

<b>Initial calibration (09.10.2019):</b>						
	<b>Date:</b>	<b>Comment:</b>				
G-Tables	4.6.02	after wedge cal. at Boulder on June 4, 2002				
N-Tables	24.4.14	after FC with D065 on April 24, 2014 with old R-N at MOHp 2014				
<b>Corrections to N-Tables</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
from SL-Test	-0,70	-0,70	-0,70	0,00	0,00	
from Comparison	-0,09	-0,04	-0,11	0,02	0,07	
Sum	-0,79	-0,74	-0,81	0,02	0,07	
<b>Comments:</b> Very good agreement in AD and acceptable in CD with moderate mue-dependance moderate mue-dependance because of larger mue-dep. in D065 C-data no FC necessary, comparisons on October 12 and 17 do not provide better data						
<b>Final Calibration (09.10.2019):</b>						
	<b>Date:</b>	<b>Comment:</b>				
G-Tables	4.6.02	after wedge cal. at Boulder on June 4, 2002				
N-Table (new)	24.4.14	old R-N-table kept				
<b>Corrections to:</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
new G-Table				0,00	0,00	
<b>Comments:</b> Final = Initial Calibration, as best results (smallest mue-dependance) on October 9, 2019, intercomparisons on October 12 and 17 do not provide better data. Old R-N-tables and SL-references kept						
<b>Reference Standard Lamp Data:</b>		<b>Date:</b> - none -				
for new R-N-table						
<b>Lamp No.</b>	<b>A</b>		<b>C</b>		<b>D</b>	
	R	N	R	N	R	N
not necessary						
<b>Q-Table:</b> Q-table checked and adjusted (if necessary) when back at station monthly check with HG-test and possibly necessary corrections as described in Komhyr's old or Bob's new Dobson manual						

## **Instrument D064 Germany (MOHp, European regional standard)**

### **Original calibration data**

N-tables from 24 April 2014, based on DSGQP-comparison with D065 on 24 April 2014 at MOHp2014, G-tables from 4 June 2002 at Boulder.  
Reference Standard Lamp Values for lamps 64Q1, 64Q2 and 64Q3.  
Lamp tests results used in data processing at home station.

### **Introductory remarks**

D064 is European Regional Reference Dobson, normally calibrated relative against World Standards D065 or D083 or absolut after Langley Method at Izana.

### **Initial/final calibration results**

(Adjustments based on the results with Standard Lamp tests included)  
9 October 2019:

**d\_Na: -0.09 d\_Nc: -0.04 d\_Nd: -0.11 d\_Nad: 0.02 d\_Ncd: 0.07**

The d\_Nad value implies an average **0.00 % error** in calculated ozone value,  $\mu=1.3$  to 2.5 (data truncated at  $\mu$ -values  $< 1.3$ ), Total Ozone = 300 Dobson Units. No  $\mu$ -dependance in AD, moderate in CD due to  $\mu$ -dependance in D065 C-data.

### **Optical, mechanical and electronical work performed**

- No more work done, only comparative measurements and regular lamp tests.

### **Final intercomparison**

See under initial/final calibration results



IC and FC on the same day, as further comparisons on 12 and 17 October do not provide better data, old R-N-table and SL-references kept, **highest difference against the standard ADDSGQP observations of D065 in  $\mu$ ue range 1.15 to 3.2 was 0.2 % in total ozone, small  $\mu$ ue-dependency. CD-observations with moderate  $\mu$ -dependence, but normally not used and probably caused by D065 C-data.**

### **Recommendations/comments**

- The results of the initial calibration are very good in AD, therefore no re-calculation necessary, especially as only AD used for data evaluation.
- Old R-N-tables kept incl. old SL-reference values.
- Correction of new MOHp-Q-tables from results of the Hg-Test immediately after return to station.
- Regular test (monthly SL and HG, at least annual Symmetry Test) and cleaning of GQP/Sundirector.

Hohenpeissenberg, 06.03.2020

### Difference N\_A-C-D to Reference Instrument D065 Initial/Final Calibration on October 9, 2019



**Intercomparison Results**  
**Irene2019 from October 7 to 18, 2019 at Irene**

**Instrument D089**  
**South Africa (Irene)**

<b>Initial calibration (09.10.2019):</b>						
	<b>Date:</b>	<b>Comment:</b>				
<b>G-Tables</b>	<b>12.9.17</b>	after wedge-cal. on Sept. 12, 2017 at ElAreno2017				
<b>N-Tables</b>	<b>13.9.17</b>	after FC on Sept. 13, 2017 with new R-G at ElAreno2017				
<b>Corrections to N-Tables</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
from SL-Test	-1.60	-1.20	-1.40	-0.20	0.20	
from Comparison	-2.51	-1.95	-1.98	-0.53	0.03	
Sum	-4.11	-3.15	-3.38	-0.73	0.23	
<b>Comments:</b> acceptable agreement in AD, data noisy, larger difference in mue-range 1.5-2 large difference in single A, C and D, no data reprocessing recommended						
<b>Final Calibration (17.10.2019):</b>						
	<b>Date:</b>	<b>Comment:</b>				
<b>G-Tables</b>	<b>13.10.19</b>	new R-G-table after wedge calibration on 13.10.19				
<b>N-Table (new)</b>	<b>17.10.19</b>	new R-N after 2. FC in 17.10.19 with new R-G				
<b>Corrections to:</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
new G-Table	-28.69	-20.72	-17.39	-11.30	-3.33	
<b>Comments:</b> Results very good in AD and used to create new R-N-tables for future data processing						
<b>Reference Standard Lamp Data:</b>		<b>Date:</b>		<b>17.10.19</b>		
for new R-N-table						
Lamp No.	A		C		D	
	R	N	R	N	R	N
89Q6	38.80	11.49	35.80	16.23	34.80	18.18
89Q7	38.70	11.38	36.00	16.46	34.90	18.29
89Q9	38.90	11.61	36.00	16.46	34.90	18.29
64Q1	39.30	12.06	36.20	16.69	35.40	18.85
<b>Q-Table:</b> no new Q-table necessary, as D089 remains on station monthly check with HG-test and possibly necessary corrections as described in Komhyr's old or Bob's new Dobson manual						

## **Instrument D089 South Africa (Irene)**

### **Original calibration data**

N-tables from 13 September 2017 based on DSGQP-comparison with D064 on 13 September 2017 at El Arenosillo, G-tables from 12 September 2017. Reference Standard Lamp Values for lamps 89Q6, 89Q7 and 89Q9. Lamp tests results used in data processing at home station.

### **Introductory remarks**

Last calibration at El Arenosillo in 2017 was repeated and confirmed.

### **Initial calibration results**

(Adjustments based on the results with Standard Lamp tests included)

9 October 2019:

**d\_Na: -2.51 d\_Nc: -1.95 d\_Nd: -1.98 d\_Nad: -0.53 d\_Ncd: 0.03**

The d\_Nad value implies an average **+0.7 % error** in calculated ozone value,  $\mu = 1.15$  to 2.5, Total Ozone = 300 Dobson Units. Measurements very noisy due to not optimal operation of the instrument, larger difference in  $\mu$ -range 1.5 - 2.0 and in single A, C and D. As AD acceptable no data reprocessing recommended.

### **Optical, mechanical and electrical work performed**

- Electric/Electronics: Old Komhyr-version, some noise problems, reason for the well-known problem of different readings (detected during SL-Test) at different setting of High Voltage found: shutter motor causes RFI on the electronics which results in an offset of the microammeter zero.
- Optical check: All optics (Ps, Ms, Ls, Q-plates, wedge dusty, surface of cobalt filter with film.
- Symmetry test: Done twice, Results satisfying in important values.
- Measurement of slit widths and parallelism with microscope: Not done.
- Shutter motor: Toothbelt driven, 815 rpm O.K.
- PMT vertical position test: and Focus L1: Not done.
- Optics: All optics cleaned, special cleaning of cobalt filter.
- Further work: Gasket replaced.
- Discharge lamp: Not done.
- Wedge calibration: Done on 12 October 2019, applied to create new R-G-tables for FC.

### **Final intercomparison**

12 and 17 October 2019

Data of two final calibration processed with the new R-G-Tables to derive new R-N-tables, much less noise in the data than during the IC; **second FC used for future R-N-tables; highest difference against the standard ADDSGQP observations of D064 in  $\mu$  range 1.15 to 3.2 was -0.7 % in total ozone, no significant  $\mu$ -dependency. CD-results O.K. with  $\mu$ -dependence, but not used.**

**Recommendations/comments**

- The results of the initial calibration are satisfactory in AD, therefore no re-calculation necessary.
- New calibration status with new R-N-tables defined based on new R-G-tables incl. determination of SL-reference values from corresponding tests (s. table with results).
- **Correction of Q-tables from results of the Hg-Test not necessary.**
- Regular test (monthly SL and HG, at least annual Symmetry Test) and cleaning of GQP/Sundirector.
- **Recommendation to avoid RFI-effect/zero-offset on SL-Test and measurements: Switch on shutter motor only with inlet window closed and correct offset of the microamperemeter if any.**
- **New absorption coefficients will officially be introduced in the near future: reprocessing of all data necessary. Information will be given. Question of the future use of effective absorption coefficients (derived from optical characterization) not finally clarified.**

Hohenpeissenberg, 28.11.2019




## Difference N\_A-C-D to Reference Instrument D064

### Initial Calibration on 9 October 2019


N(D064)-N(D089) vs. Mue for D064 & D089 Irene, 9. October 2019

Initial Calibration original data - truncated & edited



N(D064)-N(D089) vs. Mue for D064 & D089 Irene, 9. October 2019


Initial Calibration adjusted data - truncated & edited




### Difference N\_A-C-D to Reference Instrument D064


- 1. & 2. Final Calibration on 12/17 October 2019
- 2. FC used for calibration and preparing of new R-N-table

N(D064)-N( D089) vs. Mue for D064 & D089 Irene, 12. October 2019  
 1. Final Calibration adjusted data - old R-N table



N(D064)-N( D089) vs. Mue for D064 & D089 Irene, 12. October 2019  
 1. Final Calibration adjusted data - new R-G





**Intercomparison Results**  
**Iren2019 from October 7 to 18, 2019 at Irene**

**Instrument D112**  
**India (New Delhi)**

<b>Initial calibration (12.10.2019):</b>						
	<b>Date:</b>	<b>Comment:</b>				
G-Tables	???	Origine not known				
N-Tables	18.3.06	from FC on March 18, 2006 at Tsukuba				
<b>Corrections to N-Tables</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
from SL-Test	-3.10	-2.70	-2.50	-0.60	-0.20	
from Comparison	1.37	1.57	2.38	-1.01	-0.81	
Sum	-1.73	-1.13	-0.12	-1.61	-1.01	
<b>Comments:</b> Agreement in AD out of limits, noisy data, large diff. especially in D Reprocessing recommended after check of SL-test history to determine doubtful periods						
<b>Final Calibration (17.10.2019):</b>						
	<b>Date:</b>	<b>Comment:</b>				
G-Tables	16.10.19	from wedge cal. on October 16, 2019, at Irene				
N-Table (new)	17.10.19	from IC on October 16, 2019 with new R-G				
<b>Corrections to:</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
new G-Table	-8.39	-5.54	-6.60	-1.79	1.06	
<b>Comments:</b> Very good agreement in AD and single A, C and D New R-N-tables derived from this IC using new R-G-table						
<b>Reference Standard Lamp Data:</b>		<b>Date:</b>		17.10.19		
for new R-N-table						
Lamp No.	A		C		D	
	R	N	R	N	R	N
112Q2	25.70	11.78	28.30	17.54	30.80	18.59
64Q1	25.70	11.78	28.40	17.64	30.80	18.59
<b>Q-Table:</b> Check new Q-table from discharge lamp test October 15 after return by doing an HG-test and correct if necessary according the procedure described in Komhyr's old or Bob's new Dobson manual						

## **Instrument D112 India (New Delhi)**

### **Original calibration data**

N-tables from 18 March 2006 based on DSGQP-comparison with D116 on 18.3.2006, Tsukuba, origin of G-tables not known.

Reference Standard Lamp Values for lamps 112Q2.

Lamp tests results used in data processing at home station.

### **Introductory remarks**

Instrument used as national standard. HG lamp not operational, therefore D064-equipment used.

### **Initial calibration results**

(Adjustments based on the results with Standard Lamp tests included)

12 October 2019:

**d\_Na: 1.37 d\_Nc: 1.57 d\_Nd: 2.38 d\_Nad: -1.01 d\_Ncd: -0.81**

The d\_Nad value implies an average **+1.4% error** in calculated ozone value,  $\mu=1.15$  to 2.5, Total Ozone = 300 Dobson Units. **AD-difference out of limits, large difference especially in D. Reprocessing recommended after check of SL-test history to determine doubtful periods.**

### **Optical, mechanical and electrical work performed**

- Electric/Electronics: Japanese electronic type, shunt resistor of microammeter corroded.
- Optical check: All optics incl. wedge slightly dusty, film on surface of cobalt filter, mirrors with fingerprints and scratches.
- Symmetry test: Acceptable at the beginning, good after mirror replacement, optical alignment however very difficult.
- Measurement of slit widths and parallelism with microscope: Not done.
- Shutter motor: Problems with correct speed, very slow (250 rpm at the beginning of operation, 750 rpm after 10s), could be easily stopped with a finger. Repair or even replacement not possible at Irene; attempt with better lubrication.
- PMT vertical position test: Good after optical alignment after mirror replacement, Focus L1: Not.
- Optics: All optics cleaned, special cleaning of surface of cobalt filter, mirrors replaced.
- Further work: Shunt resistor and gasket replaced.
- Discharge lamp: Done twice, on 9 October before initial calibration and on 15 October after optical alignment, new Q-table created.
- Wedge calibration: Done on 16 October, manually data acquisition necessary, as encoder of wedge calibrator defective. New R-G-tables created.

### **Final intercomparison**

17 October 2019

Data of final calibration processed with the new R-G-Tables to derive new R-N-tables as results better than with old R-N-tables, very good agreement in AD, small noise in the

data; **FC used for future R-N-tables; highest difference against the standard ADDSGQP observations of D064 in mu range 1.15 to 3.2 was -0.3 % in total ozone, no significant mue-dependency. CD-results O.K. with mue-dependence, but not used.**



#### **Recommendations/comments**

- The results of the initial calibration are out of 1%-limit, therefore re-calculation recommended after check of SL-test history to determine doubtful periods.
- New calibration status with new R-N-tables defined based on new R-G-tables incl. determination of SL-reference values from corresponding tests (s. table with results).
- **Correction of new Irene-Q-tables from results of the Hg-Test immediately after return to station.**
- Regular test (monthly SL and HG, at least annual Symmetry Test) and cleaning of GQP/Sundirector.
- **New absorption coefficients will officially be introduced in the near future: reprocessing of all data necessary. Information will be given.**
- **Electronic and motor replacement recommended within the next years.**

Hohenpeissenberg, 27.12.2019

## Difference N\_A-C-D to Reference Instrument D064

### Initial Calibration on 12 October 2019




### Difference N\_A-C-D to Reference Instrument D064

### Final Calibration on 17 October 2019


N(D064)-N(D112) vs. Mue for D064 & D112 Irene, 17. October 2019

Final Calibration adjusted data - new Gtable



N(D064)-N(D112) vs. Mue for D064 & D112 Irene, 17. October 2019

Final Calibration adjusted data - old Ntable





**Intercomparison Results**  
**Iren2019 from October 7 to 18, 2019 at Irene**

**Instrument D132**  
**South Africa (Springbok)**

<b>Initial calibration (09.10.2019):</b>						
	<b>Date:</b>		<b>Comment:</b>			
G-Tables	22.10.09		after wedge-cal. on Oct. 22, 2009 at Irene2009			
N-Tables	27.10.09		after FC on Oct. 27, 2009 with new R-G at Irene2009			
<b>Corrections to N-Tables</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
from SL-Test	-3.90	-3.90	-3.50	-0.40	-0.40	
from Comparison	-0.69	0.09	-0.50	-0.19	0.59	
Sum	-4.59	-3.81	-4.00	-0.59	0.19	
<b>Comments:</b> Good agreement in AD, noisy data in A, moderate $\mu$ -dependence Used as results more consistent in single D than that of 2. IC on 12.10. No data reprocessing recommended.						
<b>Final Calibration (17.10.2019):</b>						
	<b>Date:</b>		<b>Comment:</b>			
G-Tables	22.10.09		after wedge-cal. on Oct. 22, 2009 at Irene2009			
N-Table (new)	17.10.19		after FC on Oct. 17, 2019 with old R-N at Irene2019			
<b>Corrections to:</b>						
	<b>A</b>	<b>C</b>	<b>D</b>	<b>AD</b>	<b>CD</b>	
old N-Table incl. SL-test	-4.30	-3.50	-3.70	-0.60	0.20	
<b>Comments:</b> Results very good in AD and good in single A, C and D new R-N-tables derived and recommended for future data processing.						
<b>Reference Standard Lamp Data:</b>			<b>Date:</b>		17.10.19	
for new R-N-table						
Lamp No.	A		C		D	
	R	N	R	N	R	N
132A	39.10	12.69	39.90	17.93	40.00	19.43
132C	39.00	12.59	39.40	17.42	39.90	19.33
132D	39.20	12.79	39.80	17.83	40.00	19.43
132E	38.50	12.39	39.60	17.66	39.70	19.15
<b>Q-Table:</b> Check original Q-table in use at station by doing an HG-test and correct if necessary according the procedure described in Komhyr's old or Bob's new Dobson manual						

## **Instrument D132 South Africa (Springbok)**

### **Original calibration data**

N-tables from 27 October 2009 based on DSGQP-comparison with D064/D083  
27 October 2009 at Irene, G-tables from 22 October 2009.  
Reference Standard Lamp Values for lamps 132A, 132C, 132D and 132E.  
Lamp tests results used in data processing at home station.

### **Introductory remarks**

Two initial calibrations performed on 9 and 12 October 2019. Only IC on first day used, as results of second IC especially in D not consistent (reverse sign, data noisy and doubtful).

### **Initial calibration results**

(Adjustments based on the results with Standard Lamp tests included)  
9 October 2019:

**d\_Na: -0.69 d\_Nc: 0.09 d\_Nd: -0.50 d\_Nad: -0.19 d\_Ncd: 0.59**

The d\_Nad value implies an average **0.3 error** in calculated ozone value,  $\mu = 1.15$  to 2.5, Total Ozone = 300 Dobson Units. Good agreement in AD with moderate  $\mu$ -dependence. No data reprocessing necessary using this first IC.

### **Optical, mechanical and electronical work performed**

- Electric/Electronics: New US-type MOHp modified.
- Optical check: All optics slightly dusty, cobalt filter dirty, wedge O.K.
- Symmetry test: Done on 8 and 15 October, good results.
- Measurement of slit widths and parallelism with microscope: Not done.
- Shutter motor:
- PMT vertical position test: Not done, Focus L1: Not done.
- Optics: All optics incl. cobalt filter cleaned.
- Discharge lamp: Done on 11 October 2019, not used as no difference to original Q-table.
- Wedge calibration: Not done.

### **Final intercomparison**

17 October 2019

Data of final calibration processed with old R-N-Tables to derive new R-N-tables, very good agreement in AD, small noise in the data; **FC used for future R-N-tables; highest difference against the standard ADDSGQP observations of D064 in  $\mu$  range 1.15 to 3.2 was -0.5 % in total ozone, no significant  $\mu$ -dependency. CD-results O.K. with  $\mu$ -dependence, but not used.**

### **Recommendations/comments**


- The results of the initial calibration are good although data noisy and have to be edited, therefore no re-calculation necessary, especially as only AD used for data evaluation.
- New calibration status with new R-N-tables defined based on old R-N-tables incl. determination of SL-reference values from corresponding tests (s. table with results).
- **Correction of original-Q-tables at station from results of the Hg-Test immediately after return to station.**
- Regular test (monthly SL and HG, at least annual Symmetry Test) and cleaning of GQP/Sundirector.
- **New absorption coefficients will officially be introduced in the near future: reprocessing of all data necessary. Information will be given.**

Hohenpeissenberg, 30.12.2019


## Difference N\_A-C-D to Reference Instrument D064

### 1. Initial Calibration on 9 October 2019

N(D064)-N(D132) vs. Mue for D064 & D132 Irene, 9. October 2019  
1. Initial Calibration original data - truncated & edited




N(D064)-N(D132) vs. Mue for D064 & D132 Irene, 9. October 2019  
1. Initial Calibration adjusted data - truncated & edited




### Difference N\_A-C-D to Reference Instrument D064 Final Calibration on 17 October 2019

N(D064)-N(D132) vs. Mue for D064 & D132 Irene, 17. October 2019  
Final Calibration original data - truncated & edited



N(D064)-N(D132) vs. Mue for D064 & D132 Irene, 17. October 2019  
Final Calibration adjusted data - truncated & edited



## LIST OF RECENT GAW REPORTS\*

252. Research Infrastructure Quality Assurance - System and Performance Audit of Surface Ozone, Carbon Monoxide, Methane, and Carbon Dioxide at the Global GAW Station Ushuaia, Argentina, November 2019, WCC-Empa Report No. 19/3.
251. Research Infrastructure Quality Assurance - System and Performance Audit of Surface Ozone, Carbon Monoxide, Methane, Carbon Dioxide and Nitrous Oxide at the Global GAW Station Izaña, Spain, May 2019, WCC-Empa Report No. 19/2.
250. Global Atmosphere Watch Expert Meeting Workshop on Measurement-Model Fusion for Global Total Atmospheric Deposition (MMF-GTAD), Geneva, Switzerland, 26-27 February 2019.
249. Report of the Fifth Session of the CAS Environmental Pollution and Atmospheric Chemistry Scientific Steering Committee (EPAC SSC), Geneva, Switzerland, 5-6 November 2018, 2020.
248. Twelfth Intercomparison Campaign of the Regional Brewer Calibration Center Europe, El Arenosillo Atmospheric Sounding Station, Huelva, Spain, 27 May–9 June 2017, 2019.
247. Izaña Atmospheric Research Center Activity Report 2017-2018, 2019.
246. Thirteenth Intercomparison Campaign of the Regional Brewer Calibration Center Europe (RBCC-E), Arosa Lichtklimatisches Observatorium, Switzerland, 30 July to 8 August 2018.
245. An Integrated Global Greenhouse Gas Information System (IG3IS) Science Implementation Plan, 2019.
244. Report of the 2017 Global Atmosphere Watch Symposium and Fourth Session of the CAS Environmental Pollution and Atmospheric Chemistry Scientific Steering Committee (EPAC SSC), Geneva, Switzerland, 10-13 April 2017, 2019.
243. Report of the Fifth Erythemal UV Radiometers Intercomparison, Buenos Aires, Argentina, 2019.
242. 19<sup>th</sup> WMO/IAEA Meeting on Carbon Dioxide, Other Greenhouse Gases and Related Tracers Measurement Techniques (GGMT-2017), Dübendorf, Switzerland, 27-31 August 2017, 2018.
241. SPARC/IOC/GAW Report on Long-term Ozone Trends and Uncertainties in the Stratosphere, SPARC Report No. 9, WCRP-2017/2018, GAW Report No. 241, 2018.
240. Report of the Second International UV Filter Radiometer Intercomparison UVC-II, Davos, Switzerland, 25 May-5 October 2017, 212 pp., 2018.
239. Calibration Methods of GC- $\mu$ ECD for Atmospheric SF<sub>6</sub> Measurements, 26 pp., 2018.
238. The Magnitude and Impacts of Anthropogenic Atmospheric Nitrogen Inputs to the Ocean, Reports and Studies GESAMP No. 97, 47 pp., 2018
237. Final Report of the 44<sup>th</sup> Session of GESAMP, Geneva, Switzerland, 4-7 September 2017, Reports and Studies GESAMP No. 96, 115 pp., 2018.
236. Izaña Atmospheric Research Center: Activity Report 2015-2016, 178 pp., 2017.
235. Vegetation Fire and Smoke Pollution Warning and Advisory System (VFSP-WAS): Concept Node and Expert Recommendations, 45 pp., 2018.

A full list is available at:

[http://library.wmo.int/opac/index.php?lvl=etagere\\_see&id=144#.WK2TTBiZNB](http://library.wmo.int/opac/index.php?lvl=etagere_see&id=144#.WK2TTBiZNB)

For more information, please contact:

**World Meteorological Organization**

**Science and Innovation Department**

7 bis, avenue de la Paix – P.O. Box 2300 – CH 1211 Geneva 2 – Switzerland

Tel.: +41 (0) 22 730 81 11 – Fax: +41 (0) 22 730 81 81

Email: [GAW@wmo.int](mailto:GAW@wmo.int)

**Website: <https://public.wmo.int/en/programmes/global-atmosphere-watch-programme>**